



Office of Public Affairs, Region III



Lisle, IL. 60532-4352 www.nrc.gov

No: III-23-013 May 31, 2023

Contact: Viktoria Mitlyng, 630-829-9662 Prema Chandrathil, 630-829-9663

NRC Proposes \$8,750 Civil Penalty to G2 Consulting Group

The Nuclear Regulatory Commission has <u>proposed</u> an \$8,750 fine to G2 Consulting Group LLC for two violations of NRC requirements associated with the control of NRC-regulated material.

The first violation involved the loss of a portable moisture density gauge containing radioactive material. The gauge is used for measuring the moisture content and density of soil and aggregate.

The second violation involved the failure of the company to implement appropriate procedures requiring the authorized user to log and return the gauge to a properly locked storage after use.

The NRC identified the proposed violations during a July 2021 <u>inspection</u> following the company's discovery that the gauge had been missing for six months. The NRC conducted its inspection at the company's offices in Troy and Ann Arbor, Michigan, and at temporary job sites in Troy and Canton, Michigan.

G2 Consulting <u>responded</u> to the violations documenting its corrective actions to prevent recurrence of the loss of a gauge.

The company has 30 days to pay the proposed penalty or contest the penalty in writing.